



Certificate report nr. H15-02320138

## CALIBRATION CERTIFICATE

**Instrument** Humidity and temperature transmitter HMD60YO  
**Serial number** X3220003  
**Manufacturer** Vaisala Oyj, Finland  
**Calibration date** 6th August 2002  
**Test procedure** Doc210426-A

The above instrument was calibrated by comparing the relative humidity and temperature readings to two HMP233 factory working standards. At the time of shipment, the instrument described above met its operating specifications.

The relative humidity readings of the two HMP233 factory working standards have been calibrated at the Vaisala factory by using Hygro M-3 dewpoint meter. Hygro M-3 dewpoint meter has been calibrated at the NIST laboratories by using the NIST primary standard. The temperature readings of the two HMP233 factory working standards have been calibrated at Vaisala Measurement Standards Laboratory (MSL) by using the MSL working standard traceable to the NIST. MSL has been accredited by the FINAS according to the ISO/IEC 17025 (K008).

### Calibration results

Reference humidity* % RH	Observed humidity % RH	Difference %RH	Permissible difference %RH
0.6	0.6	0.0	±2.0
38.7	38.3	-0.4	±2.0
75.9	75.9	0.0	± 2.0
Reference temperature* °C	Observed temperature °C	Difference °C	Permissible difference °C
+ 21.61	+21.67	0.06	± 0.1

\*Average of two references.

### Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
HMP233 / RH	S0120017	2002-07-11	H15-02280231
HMP233 / RH	S0120018	2002-07-11	H15-02280232
HYGRO M-3	0351095	2001-12-07	266299-02
HMP233 / T	S0120017	2001-12-12	K008-J12121
HMP233 / T	S0120018	2001-12-12	K008-J12122
HP 34401A	3146A73362	2002-07-12	K004-02S463

### Uncertainties ( 95 % confidence level, k=2)

Humidity ±1.0%RH @ 0..15%RH, ±1.5%RH @ 15..78%RH

Temperature ± 0.15 °C

Ambient conditions / Humidity 46 ± 5%RH, Temperature 22 ± 1 °C, Pressure 1018 ± 1 hPa.

For Vaisala Oyj

Anu Koivisto

This report shall not be reproduced except in full, without the written approval of Vaisala.

Doc210425-A

Vaisala Oyj  
 Vanha Nurmijärventie 21  
 FIN-01670 Vantaa, Finland

Tel. (+ 358 9) 8949 2658  
 Fax (+ 358 9) 8949 2295  
<http://www.vaisala.com>

Domicile Vantaa, Finland  
 Trade Reg. No. 96.607



Certificate report nr. H15-02330012

## CALIBRATION CERTIFICATE

**Instrument** Temperature transmitter HMD60T  
**Serial number** X3310002  
**Manufacturer** Vaisala Oyj, Finland  
**Calibration date** 12th August 2002  
**Test procedure** Doc210426-A

The above instrument was calibrated by comparing the temperature readings to two HMP233 factory working standards. At the time of shipment, the instrument described above met its operating specifications.

The temperature readings of the two HMP233 factory working standards have been calibrated at Vaisala Measurement Standards Laboratory (MSL) by using the MSL working standard traceable to the NIST. MSL has been accredited by the FINAS according to the ISO/IEC 17025 (K008).

### Calibration results

Reference temperature* °C	Observed temperature °C	Difference °C	Permissible difference °C
+ 23.74	+23.74	0.00	± 0.1

\*Average of two references.

### Equipment used in calibration


Type	Serial number	Calibration date	Certificate number
HMP233 / T	S0120017	2001-12-12	K008-J12121
HMP233 / T	S0120018	2001-12-12	K008-J12122
HP 34401A	3146A73362	2002-07-12	K004-02S463

### Uncertainties ( 95 % confidence level, k=2)

Temperature ± 0.15 °C

Ambient conditions / Humidity 42 ± 5%RH, Temperature 24 ± 1 °C, Pressure 1007 ± 1 hPa.

For Vaisala Oyj

  
 Vesa Heikkinen

*This report shall not be reproduced except in full, without the written approval of Vaisala.*


Doc210485-A

Vaisala Oyj  
 Vanha Nurmijärventie 21  
 FIN-01670 Vantaa, Finland

Tel. (+ 358 9) 8949 2658  
 Fax (+ 358 9) 8949 2295  
<http://www.vaisala.com>


Domicile Vantaa, Finland  
 Trade Reg. No. 96.607

NOTE: The ideal mounting locations are away from possible moisture sources and away from any source which can alter the ambient temperature, e.g. cooling towers. Installation on the wall is also not recommended.




Ur	Qty	Change	Material	ECO no	Design	Date Review	Date Appr
----	-----	--------	----------	--------	--------	-------------	-----------


Front View



Side View



Top View



Dimensions in mm.

		Weight	Finish	General Reference
Checker	2001-08-24 PP			Sheet 1/1
Designer	PP	<b>VAISALA</b> DTR503B RADIATION SHIELD INSTALLATION DRAWING		
Reviewer	PP			
Scale	1:4			
Approved				
		Title		Rev
		DTR503B RADIATION SHIELD INSTALLATION DRAWING		A
		Code		DRW211819

NOTE: Verify that the conduit adapter and the cover of the HMD60/70 transmitter are tightened so that they are waterproof.

## HUMIDITY TRANSMITTER HMD60U

## HUMIDITY AND TEMPERATURE TRANSMITTER HMD60Y

### MOUNTING




Figure 1 Dimensions of the HMD60U/Y

### GROUNDING





Figure 2 Signal cable grounding with bushing 18941HM

The HMD60U/Y humidity and temperature transmitters are two-wire transmitters. They are duct mounted, and the electronics can be disconnected without dismantling the installation.

Mount the transmitter with two screws. Place the drilling template on the duct surface and drill the holes as indicated. Remember to drill an additional hole for calibration purposes. The calibration can be conveniently performed on site with the HMI41 indicator equipped with an appropriate probe and optional calibration cable.

Open the lid and mount the cable bushing set 18941HM. Do the grounding according to Figure 2. When connecting the signal cable to the transmitter housing, fold the cable braid between the brass disk in order to achieve the best EMC performance. Do not leave the bare shield of the connected wires so that it can shortcircuit the electronics!


## ELECTRICAL CONNECTIONS



*Figure 3: Electrical connections*

Signal cables are connected to a removeable 5-pole screw connector. Make the connections according to Figure 3 above. RH test and T test connectors are used with the HMI41 indicator equipped with an appropriate probe and optional calibration cable.

## ELECTRONICS



### TO REMOVE THE SENSOR HEAD:

1. Open the lid
2. Disconnect the screw terminal
3. Open the screws (2 pcs)
4. Pull out carefully

### TO REINSTALL THE SENSOR HEAD:

1. Push in the sensor head
2. Fasten the screws
3. Connect the screw terminal
4. Close the lid

*Electronics (can be disconnected), accessories, spare parts*

## ONE-POINT RH-CALIBRATION

The accuracy is recommended to be checked at least once a year; the interval depends on the operating conditions and the required accuracy of the measurement. The transmitter calibration can be conveniently checked with the HMI41 indicator equipped with an appropriate probe and optional calibration cable. If adjustment is needed, use the one-point calibration potentiometer. If you prefer to calibrate the HMD60U/Y transmitters against saturated salt solutions, use LiCl (11 %RH) and NaCl (75 %RH) solutions.

## REPLACEMENT OF THE HUMICAP SENSOR AND THE FILTER

Remove the damaged sensor and insert a new one. Recalibrate the transmitter. Replace a dirty filter (membrane or sintered) to ensure a maximum lifetime and a fast response for the sensor. Do not attempt to clean the filter.


## TECHNICAL DATA

### Relative humidity


Measurement range 0...100 %RH

Output signal

Accuracy at +20 °C



Temperature dependence



Humidity sensor HUMICAP®180

Response time (90%)

at 20 °C in still air


15 s with membrane

filter

### Temperature (Y model only)

Measurement range -20...+80 °C

Accuracy



Linearity

Temperature sensor

better than 0.1 °C

Pt 1000 IEC 751 class B

### General

Supply voltage

10...35 VDC ( $R_L = 0\Omega$ )

20...35 VDC ( $R_L = 500\Omega$ )

Output signal

Operating temperature range:

electronics

-5...+55 °C

sensor head

-40...+80 °C

Storage temperature range

-40...+80 °C

Housing:

sensor head

stainless steel

electronics housing

cast aluminium

Cable lead-through:

bushing

for 7...10 mm (PG9)

cable (housing IP65 /

NEMA 4),

part no. 18941HM

part no. 10528HM

or armoured cable glands

Sensor protection:

standard

membrane filter

(part no. 17039HM)

stainless steel sintered

filter

(part no. HM46670)

screw terminals

0.5...1.5 mm<sup>2</sup>

option

Connections

### Electromagnetic compatibility

The emission and immunity tests have been performed according to standards EN50081-1 and EN50082-1.

Emissions:

Test

Setup according to

Performance

Radiated

EN55022

class B

interference

Immunity:

Test

Setup according to

Performance

Electrostatic

IEC 801-2:1991

criteria B

discharge

Electrical fast

IEC 801-4:1988

criteria B

transients

RF-radiated

IEC 801-3:1984

criteria A

fields

\*GSM-field

ENV50204:1995

criteria A

immunity

(\*additional test)



### GUARANTEE

Vaisala issues a guarantee for the material and workmanship of this product under normal operating conditions for one year from the date of delivery. Exceptional operating conditions, damage due to careless handling or misapplication will void the guarantee.